### CLICK ANYWHERE on THIS PAGE to RETURN to DISINFECTION of BUILDING SURFACES at InspectApedia.com

70-3124-15 24oz CaviCide\_3124-15 24oz CaviCide 5/19/11 2:25 PM Page 1

CAVICIDE EFFECTIVELY KILLS HUMAN IMMUNOD- FIRST AID EFICIENCY VIRUS TYPE 1 (HIV), HEPATITIS B VIRUS (HBV), AND HEPATITIS C VIRUS (HCV) ON with water for 15-20 minutes. Remove contact lenses, if PRECLEANED ENVIRONMENTAL SURFACES/ OBJECTS PREVIOUSLY SOILED WITH BLOOD/BODY FLUIDS IN HEALTHCARE SETTINGS OR OTHER SETTINGS IN WHICH THERE IS AN EXPECTED LIKELIHOOD OF SOILING OF INANI-MATE SURFACES/OBJECTS WITH BLOOD/BODY FLUIDS, AND IN WHICH THE SURFACES/ OBJECTS CAN BE ASSOCIATED WITH THE POTENTIAL FOR TRANSMISSION OF HIV-1. HBV AND HCV.

SPECIAL INSTRUCTIONS FOR CLEANING AND DECONTAMINATION AGAINST HIV-1, HBV AND HCV OF SURFACES/OBJECTS SOILED WITH BLOOD/BODY FLUID:

Personal Protection: Wear appropriate barrier protection such as protective gloves, gowns, masks and eye coverinas

ffective

 $(\otimes)$ 

Infectious Materials Disposal: Cleaning materials used that may contain blood or other bodily fluids must be autoclaved and/or disposed of in accordance with able. local regulations for infectious materials disposal.

sterilant/high-level disinfectant on any surface or instru- at 800-841-1428 or visit our website at ment that (1) is introduced directly into the human body. either into or in contact with the bloodstream or normally sterile areas of the body, or (2) contacts intact mucous membranes but which does not ordinarily penetrate the blood barrier or otherwise enter normally sterile areas of the body. This product may be used to MADE IN USA preclean or decontaminate critical or semi-critical medical devices prior to sterilization/high-level disinfection.

If in eyes: Hold eye open and rinse slowly and gently present, after the first 5 minutes, then continue rinsing eve Call a poison control center or doctor for treatment advice

If on skin: Take off contaminated clothing. Rinse skin immediately with plenty of water for 15-20 minutes. Call a poison control center or doctor for further treatment advice

For Chemical Emergency ONLY (spill, leak, fire, exposure or accident), call CHEMTREC at 1-800-424-9300. For ALL other inquiries about this product, call Metrex Customer Service at 1-800-841-1428, Monday through Friday, between 6 a.m. and 4 p.m. Pacific Time.

Have the product container or label with you when calling a poison control center or doctor, or going for treatment

STORAGE AND DISPOSAL Do not contaminate water, food, or feed by storage and disposal

Pesticide Storage: Store in a cool place. Pesticide Disposal: Wastes resulting from the use of this product may be disposed of according to applicable Federal, State, or local procedures.

Container Disposal: Nonrefillable container. Do not reuse or refill this container. Offer for recycling, if avail-

This product is not to be used as a terminal For product or technical information, contact Metrex www.metrex.com.

> Manufactured By: METREX® RESEARCH 28210 Wick Road • Romulus, Michigan 48174

Reorder No. 13-1024 70-3124-15



# decontaminant cleaner

### Bactericidal

- Virucidal\*
- Fungicidal
- Tuberculocidal\*\*

### ACTIVE INGREDIENTS:

Diisobutylphenoxyethoxyethyl dimethyl benzyl	
ammonium chloride	0.28%
Isopropanol	17.20%
INERT INGREDIENTS:	82.52%
TOTAL	100.00%

## KEEP OUT OF REACH OF CHILDREN CAUTION

#### PRECAUTIONARY STATEMENTS Hazards to Humans and Domestic Animals. CAUTION

Harmful if absorbed through skin. Causes moderate eye irritation. Avoid contact with eyes, skin or clothing. Wash hands before eating, drinking, chewing gum, using tobacco, or using the toilet. Remove contaminated clothing and wash clothing before reuse. See other panel for additional precautionary statements.

Physical or Chemical Hazards: Combustible. Do not use or store near heat or open flame.

EPA REG. NO.: 46781-6 EPA EST. NO.: 46781-MI-1

#### **EFFECTIVE AGAINST**

 Mycobacterium tuberculosis var: bovis (BCG)\*\* • Staphylococcus aureus Pseudomonas aeruginosa • Salmonella enterica • Trichophyton mentagrophytes • Methicillin Resistant Staphylococcus aureus (MRSA) • Vancomycin Resistant Enterococcus faecalis (VRE) . Staphylococcus aureus with reduced susceptibility to Vancomvcin . Hepatitis B Virus (HBV)\* . Hepatitis C Virus (HCV)\* • Herpes Simplex Virus Types 1 and 2\* • Human Immunodeficiency Virus (HIV-1)\* • Human Coronavirus (not associated with Severe Acute Respiratory Syndrome or SARS)\* • Influenza A2 Virus\*

#### DESCRIPTION

CaviCide is a multi-purpose disinfectant/decontaminant cleaner for use on hard, non-porous inanimate surfaces.

It is useful in health care settings such as hospitals, emergency medical settings, surgical centers, isolation areas, neonatal units, patient care areas, ophthalmic and optometric facilities, dental offices, dental operatories, animal care facilities, salon settings, emergency vehicles, schools, health club facilities, kitchens\*\*, bathrooms, prisons, morgues, and other critical care areas where control of cross contamination between treated surfaces is important. CaviCide will effectively clean and disinfect, when used as directed, such items as infant/child care equipment surfaces, toys, oxygen hoods, operating room tables and lights, ambulance equipment surfaces, exam tables, IV poles, dental unit instrument trays, exterior surfaces of amalgamators and dental curing lights, exterior surfaces of anesthesia machines and respiratorv therapy equipment. SilkPeel™ with DermalInfusion™. veterinary care surfaces, tanning beds, salon surfaces, exterior surfaces of appliances such as refrigerators, microwaves, food processors and stoves\*\*\*, sinks\*\*\*, counter tops\*\*\*, floors, walls, handrails, door knobs, bed railings, tables, chairs, bath-room fixtures, computers, shopping carts, cabinets, lights and telephones, which are made of plastic (i.e., polycarbonate, polypropylene, polyvinylchloride, polystyrene), and use sites listed on this label consisting of non-porous vinvl, stainless steel, painted surfaces, Plexiglas®, Formica® and glass. Kitchen\*\*\* and Bath Solution for Hard, Non-porous Surfaces Cleans, Deodorizes and Kills Common Household Germs

DIRECTIONS FOR USE

It is a violation of Federal law to use this product in a manner inconsistent with its laheling FOR PRECLEANING NON-INSTRUMENT SURFACES PRIOR TO DISIN-

FECTION: Remove gross filth and heavy soil loads. Spray CaviCide directly onto surface. Allow to remain visibly wet for 30 seconds. Wipe surface using a towel

FOR PRECLEANING INSTRUMENTS PRIOR TO DISINFECTION: Remove gross filth and heavy soil loads.

For use as a precleaning decontaminant spray: Place instruments in a

container. Spray CaviCide so as to thoroughly drench all instrument surfaces and allow to remain visibly wet for 30 seconds. Remove and rinse instruments. Follow with appropriate cleaning and disinfection process.

FOR USE AS A VIRUCIDE\* ON NON-INSTRUMENT SURFACES, INCLUD-ING HIV-1, HBV AND HCV APPLICATIONS: Spray CaviCide directly onto precleaned surface, thoroughly wetting area to be treated. Allow surface to remain visibly wet for 2 minutes at room temperature (68 °F/20°C).

FOR USE AS A DISINFECTANT ON NON-INSTRUMENT SURFACES: Spray CaviCide directly onto precleaned surface, thoroughly wetting area to be disinfected. Allow surface to remain visibly wet for 3 minutes at room tem-

perature (68 °F/20 °C) for Tuberculocidal\*\* activity and effectiveness against Staphylococcus aureus. Pseudomonas aeruginosa. Trichophyton mentagrophytes and Salmonella enterica.

Wipe surface using a towel or allow to air dry.

\*\*A thorough rinse with potable water is required for surfaces in direct con-

FOR USE AS A VIRUCIDE\* ON PRECLEANED NON-CRITICAL MEDICAL DEVICES', INSTRUMENTS AND IMPLEMENTS INCLUDING HIV-1, HBV AND HCV APPLICATIONS: Instruments must be thoroughly cleaned to remove excess organic debris, rinsed and dried. Thoroughly clean and rinse lumens of hollow instruments. Using either a tray or an ultrasonic unit, immerse instruments in **CaviCide** for 2 minutes at room temperature (68 °F/20 °C)

FOR USE AS A DISINFECTANT ON PRECLEANED NON-CRITICAL MEDICAL DEVICES'. INSTRUMENTS AND IMPLEMENTS: Instruments must be thoroughly cleaned to remove excess organic debris, rinsed and dried. Thoroughly clean and rinse lumens of hollow instruments. Using either a tray or an ultrasonic unit, immerse instruments in CaviCide for 3 minutes at room temperature (68 °F/20 °C) for Tuberculocidal\*\* activity and effectiveness against Staphylococcus aureus, Pseudomonas aeruginosa, Trichophyton mentagrophytes and Salmonella enterica.

Remove and rinse instruments. Wipe dry prior to use. Discard solution after each use.

1 Non-critical medical devices are items that come in contact only with intact skin

Note: Critical and semi-critical devices must be followed by appropriate terminal sterilization/high-level disinfection process.



Metrex

Net Contents: 24 fl. oz. (709 mL)